

Magnetic Separator - 96 Well Plate 472236

Gather Components 96-Well Plate Each well containing a Magnetic Separator suspension of magnetic particles. 1. Placement Place the 96-Well Plate on to the Magnetic Separator. 2. Empty the 96-Well Plate While firmly holding the 96-Well Plate against the Magnetic Separator, invert over a sink to discard the liquid. The magnetic particles will be retained in the 96-Well Plate. 3. Remove the 96-Well Plate The magnetic particles retained in the plate are now ready for further use. 00000000000

Gold Standard Diagnostics 124 Railroad Drive Warminster, PA 18974 WEB: www.abraxiskits.com T (215) 357 3911 F (215) 357 5232 Ordering: info.abraxis@us.goldstandarddiagnostics.com Technical Support: support.abraxis@us.goldstandarddiagnostics.com