



Please note that the ABRAXIS® Microcystins Strip Test for Recreational Water (PN 520022 and 520023) and the ABRAXIS® Microcystins Strip Test for Source Drinking Water (PN 520019 and 520020) are for the screening of **fresh water samples only**. Please contact Gold Standard Diagnostics for the *Brackish or Sea Water Sample Preparation* technical bulletin for information on the preparation and testing of brackish and sea water samples using the ABRAXIS® Microcystins Strip Test for Finished Drinking Water (PN520016 and 520017).

For ordering or technical assistance contact:

Gold Standard Diagnostics
124 Railroad Drive
Warminster, PA 18974
WEB: www.abraxiskits.com

Phone: (215) 357 3911
Fax: (215) 357 5232

Ordering: info.abraxis@us.goldstandarddiagnostics.com
Technical Support: support.abraxis@us.goldstandarddiagnostics.com

Date this Technical Bulletin is effective: 09NOV2021

Version: 01

Gold Standard Diagnostics Horsham, Inc.

124 Railroad Drive, Warminster, PA 18974 USA

(215) 357-3911 | info.abraxis@us.goldstandarddiagnostics.com | Website: www.abraxiskits.com